

CERTIFICATE OF CALIBRATION

Customer/Location:

Customer

Address Line 1

Address Line 2

City, State ZIP

USA

PO #: PO123456

Invoice #: YYMM##

PRODUCT: TC-25+ TEMPCHECK

Model: F83024-03

Serial #: 2505 4735 0033

NIST-Traceable 0.0 °C Reference Thermometer

Manufacturer: Control Company

Meter/Probe Model: 4000,90080-09

Meter/Probe S/N: 250679570

Type: Thermometer w/ thermistor probe

Calibrated: 11/12/2025

CALIBRATION DATE: 07/08/2026

CALIBRATION DUE: 07/08/2027

Temperature Range: 0.0 °C to 40.0 °C

Resolution: 0.1 °C

Accuracy: ±0.2 °C

NIST-Traceable 40.0 °C Reference Thermometer

Manufacturer: Control Company

Meter/Probe Model: 4000,90080-09

Meter/Probe S/N: 250679570

Type: Thermometer w/ thermistor probe

Calibrated: 11/12/2025

2-POINT CALIBRATION TEMPERATURES

	REFERENCE (°C)	TC-25+ (°C)	TOLERANCE (°C)
TEMPERATURE #1	0.601	0.60	±0.15
TEMPERATURE #2	39.783	39.78	±0.15

Hampshire Controls certifies that the TC-25/TC-25+ TEMPCHECK is manufactured, tested, and calibrated by factory trained personnel to meet or exceed our strict specifications and advertised accuracy and precision. Hampshire Controls uses instrumentation and standards traceable to the U.S. National Institutes of Standards and Technology (NIST) and the International Temperature Scale (ITS). See Ripple, D. (1999), NIST 60. NIST ITS-90 Thermocouple Database, World Wide Web-Internet and Web Information Systems.

Signed for and on behalf of Hampshire Controls Corporation



Michael J. Pilon, PhD

President

Name and Title

July 8, 2026

Dover, New Hampshire, United States of America

Date and Place of Issue



Hampshire Controls

Mailing Address:
PO Box 516
Dover, NH 03821

Physical Address:
1 Grove St.
Dover, NH 03820

(866) 496-9424 (toll free)
(603) 749-9424 (office)
(603) 749-9433 (fax)

www.hampshirecontrols.com

sales@hampshirecontrols.com